

PRODUCT INFORMATION LEAFLET

ACIRS-S1A-2011

Reference Material for determination of total sulfur in coal containing approximately 250 g at a nominal top-size of -212 μm .

Traceable to SI units through NIST SRM 2693 and NIST SRM 1632d.

Certified against AS 1038.6.3 series of methods for sulfur and AS1038.3 for moisture.

ACIRS-S1A-2011	Units	Total Sulfur	Standard Deviation
	%, dry basis	0.42	0.011

Samples should be thoroughly mixed by end-over-end rotation before sub-sampling. Sample bottles should be kept tightly sealed and stored in a cool, dark place.

The full technical report and Safety Data Sheet for this product are available at www.acirs.com.au/products

Date of Certification: July, 2011
 Valid to: July, 2018
 Document Number: ACIRS-S1A-2011-PIL-rev01
 Approved by: Mark Bennetts
 Australian Coal Industry Reference Samples (ACIRS)

Revision History

Document Number	Summary	Date
S1A-2011-CoA	Original in ISO Guide 31 format	14/10/2013
ACIRS-S1A-2011-PIL_rev0	Traceability statement added	02/12/2015
ACIRS-S1A-2011-PIL_rev01	Revision history and minor editorial change	27/03/2017

Australian Coal Industry Reference Samples (ACIRS)

PO Box 2315, DANGAR NSW 2309 AUSTRALIA

Phone: +61 (2) 4926 4870 Email: acpsnational@acps.com.au