

PRODUCT INFORMATION LEAFLET

ACIRS-S1A-2011

Reference Material for determination of total sulfur in coal containing approximately 250 g at a nominal top-size of -212 μm .

ACIRS-S1A-2011	Units	Total Sulfur	Standard Deviation
	%, dry basis	0.42	0.011

* Certified against AS 1038.6.3 series of methods for sulfur and AS1038.3 for moisture. Traceable to the certified values for sulfur on NIST 2693 and NIST 1632d.

Samples should be thoroughly mixed by end-over-end rotation before sub-sampling. Sample bottles should be kept tightly sealed and stored in a cool, dark place.

Date of Certification: July, 2011
Valid to: July, 2018
Approved by: Mark Bennetts

The full technical report and SDS for this product are available at
www.acirs.com.au/products

ACIRS-S1A-2011 Product Information Leaflet_rev0

Australian Coal Industry Reference Samples (ACIRS, PO Box 2315, DANGAR NSW 2309 AUSTRALIA)

Phone: +61 (2) 4926 4870

Fax: +61 (2) 4926 4902

Email: acpsnational@acps.com.au