

PRODUCT INFORMATION LEAFLET

ACIRS-S1C-2011

Reference Material for determination of total sulfur in coal containing a minimum mass of 250 g at a nominal top-size of -212 μm .

Traceable to SI units through NIST SRM 2693 and NIST SRM 1632d.

Certified against AS 1038.6.3.3 for sulfur and AS 1038.3 for moisture.

ACIRS-S1C-2011	Units	Total Sulfur	Standard Deviation
	%, dry basis	1.46	0.020

Samples should be thoroughly mixed by end-over-end rotation before sub-sampling. Sample bottles should be kept tightly sealed and stored in a cool, dark place.

The full technical report and Safety Data Sheet for this product are available at www.acirs.com.au/products

Date of Certification: December, 2011
 Valid to: December, 2019
 Document number: ACIRS-S1C-2011-PIL-rev02
 Approved by: Mark Bennetts
 Australian Coal Industry Reference Samples (ACIRS)

Revision History

Document Number	Summary	Date
S1C-2011-CoA	Original	14/10/2013
ACIRS-S1C-2011-PIL_rev0	Traceability statement added	02/12/2015
ACIRS-S1C-2011-PIL_rev01	Revision history and minor editorial change	27/03/2017
ACIRS-S1C-2011-PIL_rev02	Updated period of validity	22/02/2018